



Bid Number/बोली क्रमांक (बिड संख्या):
GEM/2024/B/5407743
Dated/दिनांक : 17-09-2024

Bid Document/ बिड दस्तावेज़

Bid Details/बिड विवरण	
Bid End Date/Time/बिड बंद होने की तारीख/समय	27-09-2024 18:00:00
Bid Opening Date/Time/बिड खुलने की तारीख/समय	27-09-2024 18:30:00
Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)	180 (Days)
Ministry/State Name/मंत्रालय/राज्य का नाम	Ministry Of Agriculture And Farmers Welfare
Department Name/विभाग का नाम	Department Of Agricultural Research And Education (dare)
Organisation Name/संगठन का नाम	Indian Council Of Agricultural Research (icar)
Office Name/कार्यालय का नाम	Icar-national Research Centre On Yak, Dirang, Arunachal Pradesh
Total Quantity/कुल मात्रा	1
Item Category/मद केटेगरी	Atomic Absorption Spectrophotometer (AAS) (Q3)
MSE Exemption for Years Of Experience/अनुभव के वर्षों से एमएसई छूट/ and Turnover/टर्नओवर के लिए एमएसई को छूट प्राप्त है	Yes
Startup Exemption for Years Of Experience/अनुभव के वर्षों से स्टार्टअप छूट/ and Turnover/ टर्नओवर के लिए स्टार्टअप को छूट प्राप्त है	Yes
Document required from seller/विक्रेता से मांगे गए दस्तावेज़	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Additional Doc 1 (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Do you want to show documents uploaded by bidders to all bidders participated in bid?/	Yes
Past Performance/विगत प्रदर्शन	10 %
Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया	No
Type of Bid/बिड का प्रकार	Two Packet Bid

Bid Details/बिड विवरण

Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय	2 Days
Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
Estimated Bid Value/अनुमानित बिड मूल्य	3200000
Evaluation Method/मूल्यांकन पद्धति	Total value wise evaluation

EMD Detail/ईएमडी विवरण

Advisory Bank/एडवाइजरी बैंक	State Bank of India
EMD Amount/ईएमडी राशि	96000

ePBG Detail/ईपीबीजी विवरण

Advisory Bank/एडवाइजरी बैंक	State Bank of India
ePBG Percentage(%) / ईपीबीजी प्रतिशत (%)	3.00
Duration of ePBG required (Months) / ईपीबीजी की अपेक्षित अवधि (महीने).	14

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy./जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज़ प्रस्तुत करने है। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable./ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए।

Beneficiary/लाभार्थी :

The Director, ICAR - NRCY

ICAR-National Research Centre On Yak, Dirang, Arunachal Pradesh, Department of Agricultural Research and Education (DARE), Indian Council of Agricultural Research (ICAR), Ministry of Agriculture and Farmers Welfare (The Director)

MII Purchase Preference/एमआईआई खरीद वरीयता

MII Purchase Preference/एमआईआई खरीद वरीयता	Yes
--	-----

MSE Purchase Preference/एमएसई खरीद वरीयता

MSE Purchase Preference/एमएसई खरीद वरीयता

Yes

1. If the bidder is a Micro or Small Enterprise as per latest definitions under MSME rules, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
2. If the bidder is a Startup, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to their meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.
3. Preference to Make In India products (For bids < 200 Crore): Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate .The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023. [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.
4. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.
5. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.
6. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 10% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Atomic Absorption Spectrophotometer (AAS) (1 pieces)

(Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively/क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक)

Technical Specifications/तकनीकी विशिष्टियाँ

* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
TYPE	Type of Atomic Absorption Spectrometer (AAS)	Double beam
	Monochromator type	Czerny-turner
	Detector Type	Photo Multiplier Tube (PMT)
SPECTRAL	Source Wavelength - Minimum	185.0 (nanometer)
	Source Wavelength - Maximum	900.0 (nanometer)
	Accuracy (+/-)	0.1 (nanometer)
	Repeatability	0.1 (nanometer)
	Diffraction grating (lines/mm)	1800
	Wavelength setting and scanning	Automatic
	Slit width Adjustment	Automatic
	Resolution (nm/mm)	0.1 (parts per million)
	Bandwidth (nm)	0.1 - 2.0
	Focal length (mm)	300 and above
LIGHT SOURCE	Type of HCL	Both Single and Multiple Element
	Number of Single element HCLs supplied (inclusive in the scope of supply)	6
	Number of multiple element (alloy) HCLs supplied (inclusive in the scope of supply)	N/A
	Measuring modes (K factor and standards)	Transmittance, Absorbance, Concentration (Wavelength, time), Emission
	Detectable Elements using single element HCL	Cu, Fe, Zn, Mn, Mg
	Detectable Elements using multi element HCL	Cu-Fe-Zn-Mn
	Measurement level	ppm level, ppb level

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Lifespan of Single element lamp (mA-Hr)	5000
	Lifespan of multi-element/Alloys lamp (mA-Hr)	5000
	Number of sockets on Turret	6
ATOMIZER	Type of atomizer	Flame Burner
	Nebulizer	Pt/Ir nebulizer
	Background Correction	D2 Lamp method
	Burner diameter (mm)	100
	Time for analysis (sample)	30.0 (second)
GENERIC	Display	through PC
	Display type	TFT
	Test result printing	Through external printer
	If inbuilt printer, Type of printer	N/A
	Safety sensors	Pressure sensor, Flame Sensor, Flame Ignition Sensor, Burner interlock, Power failure protection, Nebulizer Sensor, Auto ignition
	Burner movement (lateral, horizontal & height)	Automatic
POWER SOURCE	Power Supply	230 V A.C ± 10%, 50 Hz
OPERATING CONDITIONS	Minimum Operating Temperature	10.0 (degree Celsius)
	Maximum Operating Temperature	35.0 (degree Celsius)
	Operating Humidity (RH) (%) at 40 degree c	80.0 (percent)
ACCESSORIES	Type of gas cylinders	Acetylene
	Type of Regulators supplied	Without Heater
	Number of Regulators Supplied (inclusive in the scope of supply)	2
	Number of Acetylene Cylinders supplied (inclusive in the scope of supply)	1, 2
	Capacity of each Acetylene cylinder (Litres)	47

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	Number of Nitrous oxide Cylinders supplied (inclusive in the scope of supply)	1
	Capacity of each Nitrous oxide cylinder (Litres)	47
	Gas cylinders having PESO certificate	Yes
	Other accessories (inclusive in the scope of supply)	Gas flow control box, Oil Moisture Filters, SS Exhaust fume hood with blower, Schematic diagram and instruction manual
	Suitable software for connecting PC & Printer (inclusive in the scope of supply)	Yes
WARRANTY / AVAILABILITY OF SPARES	warranty (Year)	1.0 (year)
	Spares and Consumables inclusive in the scope of supply	Yes
CERTIFICATION	Availability of UL/CE Certification as per EN 61010-1:2010 Safety requirements for electrical equipment for measurement, control, and laboratory use General requirements	Yes
	Conformance to EMC/EMI as per EN 61326-1:2013 Electrical equipment for measurement, control and laboratory use	Yes
	Availability of Test Reports from Central Govt / NABL approved / ILAC accredited Lab to prove conformity to the specification	No

Additional Specification Parameters - Atomic Absorption Spectrophotometer (AAS) (1 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)

Specification Parameter Name	Bid Requirement (Allowed Values)
1. Required Accessories 2. PERFORMANCE GUARANTEE	1 (1). Oil Free Compressor, Fume Hood, Regulators C2H2 /Ar/Online UPS and PC and any three common element Lamp (HCL) 2 (1). Greater than 0.7 abs for 5 mg/L copper solution with an RSD of less than 0.5%. 2 (2). Instrument should have facility to check "Exact Wavelength setting "before starting the analysis which confirms the maximum energy through put for a set elemental lamp.
Additional SoW	All the safety functions must be provided as standard including Gas Leak Check; Flame Monitor; Gas Pressure Monitor; Interlock for wrong burner head use; Auto switch-over between Air-C2H2 / N2OC2H2 flame with flow rate monitoring using optical sensor; Drain Tank Level Monitor, Automatic flame extinction during power sudden power interruption, Vibration sensor for auto flame extinction; Internal fan stop sensor and flame-retardant body
1. The AAS must have future upgradation facility for the following	1. Automatic/continuous flow Hydride vapour generator with peristaltic pump 2. Graphite furnace automizer: Flame to furnace change over must be automatic. 3. All Pre installation requisites (PIR) are to be supplied by the vendor.
2. The AAS must have future upgradation facility for the following	4. Utilities required for AAS operation like Branded i3 PC , Exhaust hood assembly (1 no) , Double stage SS C2H2 Regulator (1 no) one time filled Acetylene Gases Cylinder (1 no.), Oil free air compressor. 5. Air filters for moisture, dust, oil & mist removal- 1 no
3. The AAS must have future upgradation facility for the following	6. Installation & training should be provided at our lab free of cost 7. Manufacturer or their Authorized channel partner must provide User list in India 8 Manufacturer or their Authorized channel partner/Indian Agent can participate in this tender. If Manufacturer wants to participate in this tender directly then Manufacturing certificate needs to submit for the quoted model mentioned positively otherwise their tender will be cancelled.

* Bidders offering must also comply with the additional specification parameters mentioned above.

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Tsering Wangchu Sharchokpa	790101,ICAR-NRC ON YAK, DIRANG, WEST KAMENG DISTRICT, PIN-790101	1	15

Buyer added Bid Specific Additional Specification - Drawings

S.No./क्र. सं.	Document Title	Description	Applicable/रिवर्स प्रभार के अनुसार जीएसटी i.r.o. Items

S.No./क्र.सं.	Document Title	Description	Applicable/रिवर्स प्रभार के अनुसार जीएसटी i.r.o. Items
1	ATOMIC ABSORPTION SPECTROPHOTOMETER WITH THE ACCESSORIES VIZ. OIL FREE COMPRESSOR, FUME HOOD, REGULATORS C2H2 /AR/ONLINE UPS AND PC AND ANY THREE COMMON ELEMENT LAMP (HCL) View	ATOMIC ABSORPTION SPECTROPHOTOMETER WITH THE ACCESSORIES VIZ. OIL FREE COMPRESSOR, FUME HOOD, REGULATORS C2H2 /AR/ONLINE UPS AND PC AND ANY THREE COMMON ELEMENT LAMP (HCL)	Atomic Absorption Spectrophotometer (AAS)(1)

Buyer added Bid Specific Additional Scope of Work

S.No./क्र.सं.	Document Title	Description	Applicable/रिवर्स प्रभार के अनुसार जीएसटी i.r.o. Items
1	ATOMIC ABSORPTION SPECTROPHOTOMETER WITH THE ACCESSORIES VIZ. OIL FREE COMPRESSOR, FUME HOOD, REGULATORS C2H2 /AR/ONLINE UPS AND PC AND ANY THREE COMMON ELEMENT LAMP (HCL) View	ATOMIC ABSORPTION SPECTROPHOTOMETER WITH THE ACCESSORIES VIZ. OIL FREE COMPRESSOR, FUME HOOD, REGULATORS C2H2 /AR/ONLINE UPS AND PC AND ANY THREE COMMON ELEMENT LAMP (HCL)	Atomic Absorption Spectrophotometer (AAS)(1)

The uploaded document only contains Buyer specific Additional Scope of Work and / or Drawings for the bid items added with due approval of Buyer's competent authority. Buyer has certified that these additional scope and drawings are generalized and would not lead to any restrictive bidding.

Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें

1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.

2. Scope of Supply

Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)

Disclaimer/अस्वीकरण

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्यवाई का आधार होगा।

---Thank You/धन्यवाद---